



中国国家强制性产品认证证书



证书编号: 2021322315003986

认证委托人名称: RUEGER S. A
 认证委托人地址: Chemin de Mongevon 9, Case postale 98. CH-1023
 Crissier, Switzerland
 生产者名称: RUEGER S. A
 生产者地址: Chemin de Mongevon 9, Case postale 98. CH-1023
 Crissier, Switzerland
 生产企业名称: RUEGER S. A
 生产企业地址: Chemin de Mongevon 9, Case postale 98. CH-1023
 Crissier, Switzerland
 产品名称: 温度传感器(不带变送器)
 系列、规格、型号: S01, S10, S20, S21, S22, S30, S31, S40, S41, S50, S60, S61, S62,
 S70, S80, S81, S82, S83, S96
 标准: GB 3836.1-2010、GB 3836.4-2010

上述产品符合强制性产品认证实施规则 CNCA-C23-01:2019 的要求, 特发此证。

发证日期: 2021 年 5 月 24 日 有效期至: 2026 年 5 月 23 日

首次发证日期: 2021 年 5 月 24 日

证书有效期内本证书的有效性依据发证机构的定期监督获得保持。

本证书的相关信息可通过国家认监委网站 www.cnca.gov.cn 查询



批准:



上海仪器仪表自控系统检验测试所有限公司



CERTIFICATE FOR CHINA COMPULSORY PRODUCT CERTIFICATION



CERTIFICATE NO: 2021322315003986

APPLICANT: RUEGER S.A
ADDRESS: Chemin de Mongevon 9, Case postale 98. CH-1023
 Crissier,Switzerland

MANUFACTURER: RUEGER S.A
ADDRESS: Chemin de Mongevon 9, Case postale 98. CH-1023
 Crissier,Switzerland

FACTORY: RUEGER S.A
ADDRESS: Chemin de Mongevon 9, Case postale 98. CH-1023
 Crissier,Switzerland

PRODUCTNAME: Temperature Probe (Transmitter excluded)
SERIES,SPECIFICATION,MODEL: S01,S10,S20,S21,S22,S30,S31,S40,S41,S50,S60,S61,S62,S70,
 S80,S81,S82,S83,S96
STANDARDS: GB 3836.1-2010、GB 3836.4-2010

This is to certify that the above mentioned product(s)complies with the requirements of implementation rules for compulsory certification (REFNO. CNCA-C23-01:2019).

Valid from:May 24, 2021

Valid until: May 23, 2026

Date of original certification:May 24, 2021

The validity of this certificate is subject to positive result of the regular follow up inspection by issuing certification body until the expiry date.

This certificate is available through CNCA's website: www.cnca.gov.cn



APPROVAL:

Guo AiHua



Shanghai Inspection and Testing Institute of Instruments and Automation Systems Co., Ltd.

<http://www.sitiias.com.cn>

Building 9,103 Cao Bao Road, Shanghai 200233,China

Tel: +86 21 64510844

s 0004891



中国国家强制性产品认证证书



证书编号：2021322315003986

附件

产品名称： 温度传感器(不带变送器)

型号规格： S01,S10,S20,S21,S22,S30,S31,S40,S41,S50,S60,S61,S62,S70,S80,S81,S82,S83,S96

防爆标志： Ex ia IIC T4~T6 Ga

使用环境温度范围： -55℃~+90℃

温度组别	使用环境温度范围
T6	-55℃~+55℃
T5	-55℃~+70℃
T4	-55℃~+90℃

电气参数：

最高输入电压	30V
最大输入电流	100mA
最大输入功率	0.75W
最大内部电容	280pF/m
最大内部电感	15μH/m

批 准：



上海仪器仪表自控系统检验测试所有限公司